

Tissue Copper Content Assay Kit

Note: Take two or three different samples for prediction before test.

Detection equipment: Spectrophotometer/microplate reader

Cat No: BC5565

Size: 100T/96S

Components:

Reagent I: Liquid 17 mL×1, store at 2-8°C. If any reagent precipitates, it can be dissolved in a water bath at 37°C.

Reagent II: Liquid 6 mL×1, store at 2-8°C.

Standard: Liquid 1 mL×1, 10mmol/L (10000 μmol/L) copper sulfate standard solution.

Description:

Copper (Cu) is one of the essential trace elements of human body, and is also an important part of protein and enzyme. It can exist inside and outside the red blood cells, and its main function is to assist hematopoietic, that is, to catalyze the synthesis of hemoglobin. Copper can properly promote the development of the human bone, promote the development of the human nervous system and the brain, and maintain the normal growth and development of infants and young children, therefore, the determination of copper ion content in the tissue can know whether the body is deficient in copper.

Under acidic conditions, Cu^{2+} is dissociated from ceruloplasmin and albumin and reacts with complexing agent 3, 5-dibromo-PAESA to produce a purple complex, which has a characteristic absorption peak at 580nm, and the absorbance is proportional to the concentration in a certain range, thus calculating the Cu^{2+} concentration.

Required but not provided:

Spectrophotometer/microplate reader, cryogenic centrifuge, water bath/constant temperature incubator, adjustable pipette, micro glass cuvette/96 well flat-bottom plate, mortar/homogenizer, ice and distilled water

Operation procedure:

I. Extraction of citric acid from samples

1. Tissue: according to the tissue weight (g): the volume of distilled water (mL) is 1:5-10. (It is recommended that add 1 mL of distilled water to 0.1 g tissue). Homogenate in ice bath, then centrifuge at 10000g for 10 minutes at 4°C. Take out the supernatant and put it on ice for test.

II. Determination procedure

1. Preheat spectrophotometer for 30 minutes, adjust wavelength to 580 nm and set zero with distilled water.
2. Preparation of 80μmol/L standard solution: Take 100μL of 10mmol/L standard solution and add 400 μL

- of distilled water to mix, that is, 2000 μ mol/L standard product; Then take 40 μ L 2000 μ mol/L standard product and 960 μ L distilled water to mix, that is, to prepare 80 μ mol/L standard solution.
- Preheat the Reagent I in water bath at 37°C for more than 10 minutes.
 - Add the corresponding reagent into the 1.5 mL EP tube/96 well flat-bottom plate according to the following table.

Reagent name (μ L)	Black tube (B)	Test tube (T)	Standard tube (S)
Distilled water	10	-	-
Sample	-	10	-
Standard	-	-	10
Reagent I	150	150	150
Reagent II	50	50	50

After fully mixing, leave it for 5 minutes at 37°C, measure the absorbance at 580 nm, and record it as A_B , A_T , A_S . Calculate $\Delta A_T = A_T - A_B$, $\Delta A_S = A_S - A_B$, Blank tube and standard tube only need to measure 1-2 times.

III. Calculation:

- Calculated according to fresh weight of tissue:

$$\text{The content of copper } (\mu\text{mol/g}) = \Delta A_T \div (\Delta A_S \div C_S) \times V_T \div W = 0.08 \times \Delta A_T \div \Delta A_S \div W$$

- Calculated according to the content of mitochondrial protein:

$$\text{The content of copper } (\mu\text{mol/g prot}) = \Delta A_T \div (\Delta A_S \div C_S) \times V_T \div (C_{pr} \times V_T) = 80 \times \Delta A_T \div \Delta A_S \div C_{pr}$$

C_S : Standard concentration, 80 μ mol/L;

V_T : Volume of distilled water in pre-treatment, 0.001L;

W : Sample quality, g;

C_{pr} : Protein concentration of supernatant, mg/mL.

Note:

- Test the absorbance immediately after incubation at 37°C for 5min. If the number of samples is too large, test them in batches and try to ensure that the determination is completed within 20min.
- If the measured light absorption value of the sample is greater than 0.5, it is recommended to dilute the sample with distilled water for determination, and pay attention to the simultaneous modification of the calculation formula.
- If the measured absorption value of the sample is less than 0.005 or close to the absorption value of the blank tube, the sample size can be appropriately increased, and the blank tube and standard tube also need to be adjusted accordingly.

Experimental example:

- Take 0.1g of Rat lung, add 1 mL of distilled water, grind it on ice, centrifuge supernatant and then operate according to the determination steps. Use 96 well flat-bottom plate to measure and

calculate $\Delta A_T = A_T - A_B = 0.118 - 0.089 = 0.029$, $\Delta A_S = A_S - A_B = 0.283 - 0.089 = 0.194$. Calculated according to fresh

weight of tissue:

Tissue Copper content ($\mu\text{mol/g}$) = $0.08 \times \Delta A_T \div \Delta A_S \div W = 0.120 \mu\text{mol/g}$.

2. Take 0.1010g of almond, add 1 mL of distilled water, grind it fully on ice, centrifuge supernatant and

operate according to the determination steps. Use 96 well flat-bottom plate to measure and calculate $\Delta A_T = A_T - A_B = 0.174 - 0.089 = 0.085$, $\Delta A_S = A_S - A_B = 0.283 - 0.089 = 0.194$. Calculated according to fresh weight of tissue:

Tissue Copper content ($\mu\text{mol/g}$) = $0.08 \times \Delta A_T \div \Delta A_S \div W = 0.347 \mu\text{mol/g}$.

3. Take 0.1053g of soybean powder, add 1 mL of distilled water, grind it fully on ice, centrifuge supernatant and operate according to the determination steps. Use 96 well flat-bottom plate to measure and calculate $\Delta A_T = A_T - A_B = 0.242 - 0.089 = 0.153$, $\Delta A_S = A_S - A_B = 0.283 - 0.089 = 0.194$. Calculated according to fresh weight of tissue:

Tissue Copper content ($\mu\text{mol/g}$) = $0.08 \times \Delta A_T \div \Delta A_S \div W = 0.599 \mu\text{mol/g}$.

Related Products:

BC5410/BC5415	Ferrous ion Content Assay Kit
BC4350/BC4355	Tissue Iron Content Assay Kit

